Single-Phase Power Meter

7110/7120

Features

- AC / DC Dual amp/watt-hour meter
- Wide range 0.001W-16KW
- Connection software attached
- Standby Power D.P.I. of 0.001W
- With crest factor ratio display
- Up to the 50 levels harmonic wave analysis capability
- 1000 sets of measurement data storage space
- Current crest factor is highest CF9





CE RS-232™ GPIB™

Accessories / Fixtures

Standard

- Power Cord
- RS232cable
- F71201 TEST BOX
- TL218 Alligator Clips
- TL208 2mm Test Probe

F71201 TEST BOX





Optional

- GPIB cable

TL218 Alligator Clips



Specification

Model Name	7110	7120
Frequency Measurement Mode	To achieved stable base frequency measurement (variation less than 1%) by voltage or current (non-inverter)	
Frequency Range	DC15Hz-10kHz	DC15Hz - 100kHz
Data Length	Dual 4096×16 RAM for voltage & current	
ADC Resolution	16 bits	
Sampling Rate	AC 50Hz/60Hz basic sampling rate 100 KSPS / 120 KSPS	
Arithmetic Precision	Watt/VRMS/IRMS/MEAN/PF/Deg/Line filter 32bits	
Frequency Filter	500Hz cut off, digital chip filter based on 25MHz	
Signal Filter	500Hz-3db digital filter based on Butterworth 50Hz-0.03% reading, 60Hz-0.05% reading	
Frequency Acquisition Mode	Voltage / current 100MHz baseband digital dynamic meter chip	
Phase Lead Detection	Subject to the current, analog/digital hybrid detecting (error less than 5 degrees)	

Range

Current (fixed/auto)	0.01A, 0.03A, 0.1A, 0.3A, 1A, 3A, 10A, 20A
Voltage (fixed/auto)	10V, 30V, 100V, 300V, 600V

Specification

	7440	7400
Model Name	7110	7120
Power Supply	Voltage 100 ~ 240Vac Frequency 50/60Hz	
Display	Seven-segment display	
Interface	RS-232	RS-232+GPIB
Flash Memory	6 Sets	
Environment	Temperature: 23°C±5°C, Humidity: 20-80%RH	
Dimension (W*H*D)	227×101×300 mm	
Weight	1.85kg	
Measurement bandwidth	DC 15Hz-10kHz	DC 15Hz - 100kHz
Harmonic (option)	Yes/NA	Yes/NA
Model	7110-10k-HARM 7110-10k	7120-100k-HARM 7120-100k
Fixture	F71201 TEST BOX	

Harmonics

Analysis base	To achieved stable fundamental frequency analysis by voltage or current (non-inverter)
Frequency Range	45Hz - 440Hz
FFT Data Length	1024
FFT Data Format	32 bits
Measurement Projects	1-50 THD, 1-50 level voltage and current V [n], A [n] 1-50 level voltage and current distortion percentage V [n%], A [n%] 1-50 level watts W [n] 1-50 level watts distortion percentage Watt W [n%] 1-50 level voltage and current angle DEG [n] Vrms, Irms, Watt, PF

Parameters Measurement Range

Vrms	0.1V-600V	PF	0.000-±1.000
Vdc	0.1V-600V	Deg	-180°-+180°
Irms	0.1mA-20A	THD	0.01%-999.99%
Idc	0.1mA-20A	Hz	15Hz-100kHz
W	0.01W-16kW		

RMS/MEAN Mode Voltage & Current Accuracy (23°C ±5°C)

15Hz ≤ f < 45Hz	±(0.1% of reading + 0.4% of range)
45Hz ≤ f ≤ 66Hz	±(0.1% of reading + 0.1 % of range)
66Hz < f ≤ 1kHz	±(0.1% of reading + 0.2 % of range)
1kHz < f≤ 10kHz	±(0.07*f % of reading + 0.3% of range)
10kHz < f ≤ 100kHz	$\pm (0.5\% \text{ of reading} + 0.5\% \text{ of range}) \pm [\{0.04 \times (\text{f-}10)\}\% \text{ of reading}]$

F unit is 1KHz

When the L-FILTER sets as ON: 45Hz $^{\sim}66$ Hz frequency range allowable error-0.03 $\%^{\sim}$ -0.05 of reading

When the AC is measured, if the fundamental frequency exceeds 200Hz, the F-Filter is required to be turned off in order to measure the most accurate value

**When the frequency range is more than 10KHz, the 7120 starts to support

DC Mode Voltage & Current Accuracy (23°C ±5°C)

10V-600V	±0.2% reading ±0.2% of range	0.01A - 20A	±{(0.2)% of reading + 0.2 % of range} ±offset
To add up the OFFSET errors of various files during measuring the DC current			

Power (W) Accuracy (23°C ±5°C)

AC power ranges (Auto or Manual)(40 ranges) range up to 16KW

Maximum Power (W) value is determined by the highest range of voltage profile

DC ±0.2% reading ±0.5% of range

 $15Hz \le f < 45Hz \pm (0.3\% \text{ of reading} + 0.2 \% \text{ of range})$

 $45Hz \le f \le 66Hz \pm (0.1\% \text{ of reading} + 0.1 \% \text{ of range})$

66Hz $< f \le 1$ kHz $\pm (0.2\% \text{ of reading } + 0.2 \% \text{ of range})$

1kHz < f \leq 10kHz \pm (0.4% of reading + 0.3 % of range) \pm [{0.06×(f)}% of reading]

 $10kHz < f \le 100kHz \pm (0.5 \% \text{ of reading} + 0.5 \% \text{ of range}) \pm [\{0.09 \times (f-10)\}\% \text{ of reading}]$

Signal Filter Error (AC) Frequency between 45-66Hz: Add 0.3% of reading. Frequency between45-66Hz: Add 1% of reading beyond CF9 Error (DC) Add range tolerance * 3

Accuracy Effect of the Phase Error of the Power

When the power factor PF is 0, the error range of Watt is

Situation 1: for 45Hz $\,<\,$ f, Add \pm 1.0% of VA

Situation 2: for 45Hz > f or f > 66Hz

Add \pm {(3.5 + 0.5×f)% of VA} for up to 100kHz as reference data

The unit for frequency f is kHz.

When the power factor is 0 < PF≤error range

When $0 < PF \le 1$ (θ : phase angle of the voltage and current)

for 45Hz \leq f \leq 66Hz. Add \pm power reading * $\{tan(\theta)^*(0.5)\}\%$

for f < 45Hz, f > 66Hz. Add ±power reading *{ $tan\theta^*(0.5\times f+0.2)$ }%

Error within 12 months Add ±(0.5% of reading)